



HAL
open science

Paraffin-Inert Atmospheric Solid Analysis Probe: A Fast and Easy Approach To Characterize Extremely Air-Sensitive Organometallic Complexes by Mass Spectrometry

Ahmad Naïm, Mathilde Farenc, Marie Hubert-Roux, Thierry Chavagnan, Virginie Cirriez, Alexandre Welle, Aurélien Vantomme, Evgueni Kirillov, Jean-François Carpentier, Carlos Afonso, et al.

► To cite this version:

Ahmad Naïm, Mathilde Farenc, Marie Hubert-Roux, Thierry Chavagnan, Virginie Cirriez, et al.. Paraffin-Inert Atmospheric Solid Analysis Probe: A Fast and Easy Approach To Characterize Extremely Air-Sensitive Organometallic Complexes by Mass Spectrometry. *Analytical Chemistry*, 2020, 92 (4), pp.2922-2925. 10.1021/acs.analchem.9b04478 . hal-02430146

HAL Id: hal-02430146

<https://normandie-univ.hal.science/hal-02430146v1>

Submitted on 13 Feb 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Paraffin-Inert Atmospheric Solid Analysis probe (piASAP): A Fast and Easy Approach to Characterize Extremely Air-Sensitive Organometallic Complexes by Mass Spectrometry

Ahmad Naim,^{†,‡} Mathilde Farenc,^{†,§} Marie Hubert-Roux,^{‡,§} Thierry Chavagnan,[‡] Virginie Cirriez,^{||} Alexandre Welle,^{||} Aurelien Vantomme,^{||} Evgueni Kirillov,[‡] Jean-François Carpentier,[‡] Carlos Afonso,^{*,‡,§} and Pierre Giusti^{*,†,§}


[†]Total Research and Technologies Gonfreville BP 27, 76700 Harfleur, France

[‡]Normandy University, INSA Rouen, UMR 6014 University of Rouen, Chimie Organique et Bioorganique - Réactivité et Analyse (COBRA), 76821 Mont Saint Aignan, France

[§]TOTAL RC - CNRS Joint Laboratory C2MC: Complex Matrices Molecular Characterization, University of Pau, University of Rouen, CNRS, 64053 Pau, France

^{||}Total Research and Technologies Feluy, Zone Industrielle C B-7181 Feluy, Belgium

[‡]University of Rennes, CNRS, ISCR (Institut des Sciences Chimiques de Rennes) UMR 6226, 35000 Rennes, France



ABSTRACT: Rational characterization of most organometallic compounds is hampered by their high reactivity in particular toward oxygen and water. Mass spectrometry experiments require to physically introduce the sample in the ionization source. So, the main challenge is to transfer air-sensitive organometallic compounds from inert atmosphere to the ionization source. In this aim, we have developed an easy technique that allows the analysis of air-sensitive compounds using the atmospheric solid analysis probe (ASAP). This method consists of a glass capillary filled with the sample (solid or liquid) and sealed by a paraffin plug to maintain the sample inert until the ionization process. It is illustrated through the structural characterization of a new highly air-sensitive dinuclear zirconium complex supported by an original switchable stilbene platform.

Early transition metallocene complexes have attracted a huge interest both in academy and industry as highly efficient α -olefin polymerization (pre)catalysts. Some of these complexes are nowadays commonly employed to obtain a specific stereochemistry, high molecular weight, narrow molecular weight distribution polymers and/or incorporation of comonomers,¹ thus ensuring the fabrication of polyolefin materials with controlled microstructure, crystallization behavior, rheological and mechanical properties.² And, all these peculiar properties of the polyolefin are largely correlated to the structure of the metallocene.³⁻⁵ However, these complexes are not easy to handle since they are quite sensitive to moisture and oxygen and are therefore difficult to characterize.⁶ NMR spectroscopy is often used, but this technique requires to be

insightful, samples already purified to a significant extent and is limited to paramagnetic metal complexes.⁷ Conversely, mass spectrometry (MS) techniques can be used to identify and separate complex mixtures of metallic complexes with no limitation in their electronic configuration states, in particular owing to their characteristic isotopic patterns.⁸⁻¹⁵ However, MS requires that the sample is introduced into the ionization source without being altered, that is without contact with air.¹⁶⁻¹⁸ The main challenge is therefore to transfer the sensitive compounds from the inert atmosphere where they are stored (*e.g.* a glovebox) to the mass spectrometer ionization source. Many developments have been made to allow the analysis of sensitive early transition organometallic compounds by MS.¹⁹⁻²² For instance, olefin polymerization reactions mediated by group 4 metal complexes have been monitored using electrospray

ionization (ESI) tandem MS.⁸ Also, MS was proved to be efficient for the analysis of highly air-sensitive organolanthanides by Evans *et al.* in 2000.²³ The authors succeeded to transfer a syringe filled with sample to the mass spectrometer in a rubber-septum-capped Schlenk flask under nitrogen atmosphere. However, the ESI source was not efficient to ionize so-called “ESMS silent ligands” such as allyl [C₃H₅]⁻ and cyclooctatetraenide [C₈H₈]²⁻ groups. To overcome this problem, Fogg *et al.* developed a MALDI target plate in the glove box directly connected to the ionization source to analyze air-sensitive compounds by coupling a MALDI-TOF mass spectrometer to an inert atmosphere glovebox through an open loading chamber.²⁴ Similarly, Lubben *et al.* have also coupled ESI to a glovebox.¹⁶ The sample infusion was realized using a syringe pump inside a glovebox connected to a ESI source through a capillary tube passing through a feedthrough of glovebox. Other atmospheric pressure ionization techniques have also been applied such as photoionization²⁵ and chemical ionization²⁶ to either characterize air-sensitive complexes, or to monitor reaction mechanisms.^{9,17,27–32} Recently, a modified mass spectrometer has been developed by McIndoe and coworkers to characterize metallocene complexes using electron ionization MS.³³ Hence, all direct analysis are based on the close proximity or coupling of a glove box to the mass spectrometer or modification of mass spectrometers.

EXPERIMENTAL SECTION

Materials. The melting point tubes (Length: 100 mm, Ø ext.: 1.55 mm, Ø int.: 1.15 mm) used in all experiments was provided from VWR supplier. The white paraffin pastilles that was used in the preparation experiments meets analytical specification of Ph. Eur. with mp 43–95 °C and it was ordered from Sigma-Aldrich.


Sample preparation. In glove box, a few tens to hundred milligrams of solid sample was dissolved in 2 mL of anhydrous toluene and the solution was transferred directly into a glass capillary sealed at one end. Then, the other end was closed by a paraffin plug. The sealed glass capillary was transported to the ASAP source. Anhydrous toluene was prepared through distillation over NaK Alloy under inert atmosphere and then was preserved over 1cm activated molecular sieves to ensure the dryness of the solvent. All preparations were performed under inert and controlled atmosphere ([O₂] < 2ppm and [H₂O] < 2ppm) in glovebox (GP Campus, JACOMEX). It is worthy to mention that no issues were encountered with solvents such as heptane, methanol and THF that have similar viscosity as toluene. The synthesis of complex E-(o)-{BisInd}(Zr(NMe₂)₃)₂ (1) and {Ind}Zr(NMe₂)₃ (2) has previously been reported.³⁴

Instrumentation. All data acquisitions were performed on hybrid quadrupole time of flight mass spectrometer (Synapt G2, Waters Corp. Wilmslow, UK) in positive atmospheric solid analysis probe (ASAP) mode (Corona voltage 4.0 kV, probe temperature 650 °C, source temperature 140 °C, desolvation gas flow 1200 L/hr). This instrument consists of a hybrid quadrupole/time-of-flight mass spectrometer, which incorporates a travelling wave (T-Wave)-based mobility separation device. Instrument control and data acquisition were carried out using MassLynx (version 4.1). The instrument is described in detail elsewhere.³⁵ The time of flight (TOF) mass spectrometer was externally *m/z* calibrated using sodium formate solution before sample analysis. The treatment of the data and preparation of figures was done using Waters

MassLynx (version 4.1) and Waters DriftScope (version 2.8) softwares.

RESULTS AND DISCUSSION

In this work, we describe a simple technique for sample introduction that is based on the atmospheric solid analysis probe (ASAP) ionization source coupled to an ion-mobility-mass spectrometer (Waters, Synapt G2). This latter ionization mode has been established by McEwen in 2005 and allows analysis of pure or solution samples directly deposited on a glass tip, where the analyte is desorbed by a hot nitrogen flux and ionized by nitrogen plasma generated by corona discharge.³⁶ The ASAP source has been used in many applications for both targeted analysis^{37–39} and untargeted analysis.^{40–42} Unlike other techniques that require uncontrolled multi sophisticated preparation steps or modified setups,^{43,44} in our new simple approach, referred to as paraffin-inert ASAP (piASAP),⁴⁵ a capillary tube containing air-sensitive compounds (liquid or solid) is sealed with paraffin plug under inert atmosphere condition. The sample is thus protected in a sealed glass capillary under inert atmosphere. And, it can be transported and carried away as long as the paraffin is intact. Herein, we describe the application of this new piASAP approach to the structural characterization of new highly air-sensitive dinuclear zirconium complex E-(o)-{BisInd}(Zr(NMe₂)₃)₂ (1) supported by an original switchable stilbene platform and analogous mononuclear zirconium complex {Ind}Zr(NMe₂)₃ (2) (Scheme 1).



Scheme 1. Structures of dinuclear E-(o)-{BisInd}(Zr(NMe₂)₃)₂ isomer (1) and mononuclear {Ind}Zr(NMe₂)₃ (2) zirconium complexes used in this study.

An ASAP source was used to ionize the sample. The principle of the source and the capillary is shown in Figure 1. The glass capillary is directly introduced into the source thanks to the ASAP probe. Paraffin is a mixture of alkanes that do not ionize well in comparison to most analytes using ASAP conditions but melts quickly, releasing the samples in the source. And, the sample is released inside the source and ionized by a nitrogen plasma induced by hot nitrogen flow and corona discharge.³⁸ The sample can be introduced as pure (solid or liquid) or in solution, as the ASAP ionization allows to use any type of solvent. However, it is necessary to keep inert atmosphere between the solvent and the paraffin plug to not dissolve it and oxidize the sample before analysis.




Figure 1. (a) paraffin-inert ASAP glass capillary. (b) paraffin-inert atmospheric solid analysis probe (piASAP) principle.

After analysis, the glass capillary can be removed slowly from the source to avoid contamination of the probe guide by the sample. If done properly, no carryover of the signal is observed. The total ion current obtained is relatively high and lasts for 2 minutes. This signal time allows to perform collision induce dissociation, ion mobility or any other post ionization experiments. Both metallocenes **1** and **2** were analyzed using piASAP technique. The piASAP spectra of **1** (Figure 2) and **2** (Figure S1) show the presence of protonated species of the molecular complex $[M+H]^+$ due to the presence of amine ($-NMe_2$) groups in both complexes. The accurate masses and isotopic distributions of complexes **1** and **2** (m/z 879.3233 and 428.1622, respectively) are in very good agreement (<5 ppm) with those expected theoretically based on the corresponding ions molecular formula (m/z 879.3208 and 428.1643, respectively). It is noteworthy to mention that even in case of low ionization efficiency analyte, the m/z values from the alkanes can be easily evidenced by their high mass defect by using high resolution mass spectrometry and also by using ion mobility, as alkanes appear at higher drift time. The control spectra and data extracted from drift time vs m/z plot where one can easily separate the paraffin and analyte ions (complex and its related ions) from the 2D map are shown in Figures S2, S3 and S4.

The ASAP experiments that were performed under air for **1** and **2** show the absence of $[M+H]^+$ ion and its related isotopic fragments. This observation reflects fast oxidation/hydrolysis of complexes **1** and **2** in air. Since the sampling cone voltage could have an effect on the gas-phase fragmentations and transmission of ions inside the source, we performed piASAP analysis at different sampling cone voltages. As shown in Table S1, Figure S5 and Figure S6, the highest relative intensity of $[M+H]^+$ ion for **1** and **2** was achieved at 40 and 10 V, respectively. This indicates that **1** has an enhanced ion transmission at 40 V in comparison to 10 V. With **2**, we clearly noticed that the increase of sampling voltage leads to more fragmentations and disappearance of isotopic distribution of $[M+H]^+$ ion at higher voltage (Figure S6). Furthermore, both structures of **1** and **2** were validated by tandem mass spectrometry experiments (MS/MS) using the piASAP technique. The piASAP MS/MS spectra of **1** and **2** (Figure S9 and Figure S10) display similar mode of fragmentation with successive loss of dimethyl amine moieties for **1** leading to the ions observed at m/z 834.2622, 790.2046 and 525.117; the ion

at m/z 613.2177 corresponds to loss of a zirconium dimethyl amine group ($Zr(N_2C_4H_{11})$).




Figure 2. Analysis of *E*-Zr dimer (**1**) by piASAP and ASAP techniques. (a) and (b) are mass spectra of piASAP and ASAP conducted under inert and aerobic conditions, respectively. The zoomed areas showcase the theoretical and experimental isotopic patterns of complex **1**.

Similarly, complex **2** shows subsequent loss of dimethyl amine groups leading to the ions observed at m/z 401.1157, 357.0669, 330.0197 and with isotopic distributions related to the gain of water molecules. This did occur because of the presence of minute amounts of water in the instrument collision cell. It is noteworthy that the detection of free ligand in piASAP MS analysis and its absence in both MSMS spectra of **1** and **2** imply that the detected free ligand is linked to the degradation of the complexes inside the source and not to collisional activation during the ions transfer.^{46,47}

CONCLUSIONS

In this study, we have described a technique referred as piASAP that allows the easy handling of extremely air sensitive solid or liquid complexes from a glove box to a mass spectrometer without modification of the instrument. We characterized effectively two metallocenes **1** and **2** using this technique by ASAP source. Both structures of **1** and **2** were confirmed by piASAP MSMS experiments and they display similar mode of fragmentation with the successive loss of amine groups. We believe that this technique will open the way for many researchers to characterize many highly air sensitive complexes in easy and fast way.

ASSOCIATED CONTENT

Supporting Information

The Supporting Information is available free of charge on the ACS Publications website at DOI: XXX

Experimental details, supporting tables, and supporting figures (PDF)

AUTHOR INFORMATION

Corresponding Author

*Tel: (33) 2 35 52 29 40. Email: carlos.afonso@univ-rouen.fr (C. A.).

*Tel: (33) 2 35 55 14 32. Email: pierre.giusti@total.com (P. G.)

ORCID

Ahmad Naim: 0000-0002-0746-6335

Mathilde Farenc: 0000-0002-5090-832X

Marie Hubert-Roux: 0000-0003-4986-4590

Evgueni Kirillov: 0000-0002-5067-480X

Jean-François Carpentier: 0000-0002-9160-7662

Carlos Afonso: 0000-0002-2406-5664

Pierre Giusti: 0000-0002-9569-3158

Author Contributions

All authors have given approval to the final version of the manuscript.

Notes

The authors declare no competing financial interest.

ACKNOWLEDGMENT

This work was supported by Total S. A. and Total Research and Technologies Feluy and Gonfreville (Postdoctoral grants for A.N. and T.C.). The authors gratefully acknowledge European Regional Development Fund (ERDF, N° HN0001343), Labex SynOrg (ANR-11-LABX-0029) and Région Normandie for their financial support.

REFERENCES

- Morse, P. M. The Lure of Single-Site Catalysts. *Chem. Eng. News Archive* **1998**, *76* (49), 25–28. <https://doi.org/10.1021/cen-v076n049.p025>.
- Kaminsky, W. *Polyolefins: 50 Years after Ziegler and Natta II: Polyolefins by Metallocenes and Other Single-Site Catalysts*; Advances in Polymer Science; Springer Berlin Heidelberg, 2013.
- Spaleck, W.; Antberg, M.; Rohrmann, J.; Winter, A.; Bachmann, B.; Kiprof, P.; Behm, J.; Herrmann, W. A. High Molecular Weight Polypropylene through Specifically DesigneZirconocene Catalysts. *Angew. Chem. Int. Ed. Engl.* **1992**, *31* (10), 1347–1350. <https://doi.org/10.1002/anie.199213471>.
- Hlatky, G. G. Metallocene Catalysts for Olefin Polymerization: Annual Review for 1996. *Coord. Chem. Rev.* **1999**, *181* (1), 243–296. [https://doi.org/10.1016/S0010-8545\(98\)00220-3](https://doi.org/10.1016/S0010-8545(98)00220-3).
- Brintzinger, H. H.; Fischer, D.; Mülhaupt, R.; Rieger, B.; Waymouth, R. M. Stereospecific Olefin Polymerization with Chiral Metallocene Catalysts. *Angew. Chem. Int. Ed. Engl.* **1995**, *34* (11), 1143–1170. <https://doi.org/10.1002/anie.199511431>.
- Henderson, W.; McIndoe, J. S. *Mass Spectrometry of Inorganic and Organometallic Compounds: Tools - Techniques - Tips*; Inorganic Chemistry: A Textbook Series; Wiley, 2005.
- Schnee, G.; Farenc, M.; Bitard, L.; Vantomme, A.; Welle, A.; Brusson, J.-M.; Afonso, C.; Giusti, P.; Carpentier, J.-F.; Kirillov, E. Synthesis, APPI Mass-Spectrometric Characterization, and Polymerization Studies of Group 4 Dinuclear Bis(Ansa-Metallocene) Complexes. *Catalysts* **2018**, *8*, 558. <https://doi.org/10.3390/catal8110558>.
- Feichtinger, D.; Plattner, D. A.; Chen, P. Ziegler–Natta-like Olefin Oligomerization by Alkylzirconocene Cations in an Electrospray Ionization Tandem Mass Spectrometer. *J. Am. Chem. Soc.* **1998**, *120* (28), 7125–7126. <https://doi.org/10.1021/ja980004s>.
- Trefz, T. K.; Henderson, M. A.; Linnolahti, M.; Collins, S.; McIndoe, J. S. Mass Spectrometric Characterization of Methylaluminum-Activated Metallocene Complexes. *Chem.: Eur. J.* **2015**, *21* (7), 2980–2991. <https://doi.org/10.1002/chem.201405319>.
- Strauch, J. W.; Fauré, Jean-Luc; Bredeau, S.; Wang, C.; Kehr, G.; Fröhlich, R.; Luftmann, H.; Erker, G. (Butadiene)Metallocene/B(C₆F₅)₃ Pathway to Catalyst Systems for Stereoselective Methyl Methacrylate Polymerization: Evidence for an Anion Dependent Metallocene Catalyzed Polymerization Process. *J. Am. Chem. Soc.* **2004**, *126* (7), 2089–2104. <https://doi.org/10.1021/ja035739y>.
- Schrader, W.; Handayani, P. P.; Zhou, J.; List, B. Characterization of Key Intermediates in a Complex Organocatalytic Cascade Reaction Using Mass Spectrometry. *Angew. Chem. Int. Ed.* **2009**, *48* (8), 1463–1466. <https://doi.org/10.1002/anie.200804353>.
- Wassenaar, J.; Jansen, E.; van Zeist, W.-J.; Bickelhaupt, F. M.; Siegler, M. A.; Spek, A. L.; Reek, J. N. H. Catalyst Selection Based on Intermediate Stability Measured by Mass Spectrometry. *Nat. Chem.* **2010**, *2*, 417.
- Iacobucci, C.; Reale, S.; De Angelis, F. Elusive Reaction Intermediates in Solution Explored by ESI-MS: Reverse Periscope for Mechanistic Investigations. *Angew. Chem. Int. Ed.* **2016**, *55* (9), 2980–2993. <https://doi.org/10.1002/anie.201507088>.
- Benigni, P.; Bravo, C.; Quirke, J. M. E.; DeBord, J. D.; Mebel, A. M.; Fernandez-Lima, F. Analysis of Geologically Relevant Metal Porphyrins Using Trapped Ion Mobility Spectrometry–Mass Spectrometry and Theoretical Calculations. *Energy Fuels* **2016**, *30* (12), 10341–10347. <https://doi.org/10.1021/acs.energyfuels.6b02388>.
- Giraldo-Dávila, D.; Chacón-Patiño, M. L.; Ramirez-Pradilla, J. S.; Blanco-Tirado, C.; Combariza, M. Y. Selective Ionization by Electron-Transfer MALDI-MS of Vanadyl Porphyrins from Crude Oils. *Fuel* **2018**, *226*, 103–111. <https://doi.org/10.1016/j.fuel.2018.04.016>.
- Lubben, A. T.; McIndoe, J. S.; Weller, A. S. Coupling an Electrospray Ionization Mass Spectrometer with a Glovebox: A Straightforward, Powerful, and Convenient Combination for Analysis of Air-Sensitive Organometallics. *Organometallics* **2008**, *27* (13), 3303–3306. <https://doi.org/10.1021/om800164e>.
- Trefz, T. K.; Henderson, M. A.; Wang, M. Y.; Collins, S.; McIndoe, J. S. Mass Spectrometric Characterization of Methylaluminum-Activated Metallocenes. *Organometallics* **2013**, *32* (11), 3149–3152. <https://doi.org/10.1021/om400256f>.
- Vikse, K. L.; Woods, M. P.; McIndoe, J. S. Pressurized Sample Infusion for the Continuous Analysis of Air- And Moisture-Sensitive Reactions Using Electrospray Ionization Mass Spectrometry. *Organometallics* **2010**, *29* (23), 6615–6618. <https://doi.org/10.1021/om1008082>.
- Zoorob, G.; Byrde Brown, F.; Caruso, J. Electrospray Mass Spectrometry as a Technique for the Elemental Analysis of Metals and Organometals. *J. Anal. At. Spectrom.* **1997**, *12* (5), 517–524. <https://doi.org/10.1039/A602682F>.
- Løver, T.; Henderson, W.; Bowmaker, G. A.; M. Seakins, J.; Cooney, R. P. Electrospray Mass Spectrometry of Highly Moisture-Sensitive Metal Alkoxides. *J. Mater. Chem.* **1997**, *7* (8), 1553–1558. <https://doi.org/10.1039/A701568B>.
- Colton, R.; D'Agostino, A.; Traeger, J. C. Electrospray Mass Spectrometry Applied to Inorganic and Organometallic Chemistry. *Mass Spectrometry Reviews* **1995**, *14* (2), 79–106. <https://doi.org/10.1002/mas.1280140203>.

- (22) Colton, R.; Dakternieks, D. An Electrospray Mass Spectrometric Study of Some Mercury Phosphine Complexes. *Inorganica Chimica Acta* **1993**, *208* (2), 173–177. [https://doi.org/10.1016/S0020-1693\(00\)85118-X](https://doi.org/10.1016/S0020-1693(00)85118-X).
- (23) Evans, W. J.; Johnston, M. A.; Fujimoto, C. H.; Greaves, J. Utility of Electrospray Mass Spectrometry for the Characterization of Air-Sensitive Organolanthanides and Related Species. *Organometallics* **2000**, *19* (21), 4258–4265. <https://doi.org/10.1021/om000178u>.
- (24) Eelman, M. D.; Blacquiere, J. M.; Moriarty, M. M.; Fogg, D. E. Shining New Light on an Old Problem: Retooling MALDI Mass Spectrometry for Organotransition-Metal Catalysis. *Angew. Chem. Int. Ed.* **2008**, *47* (2), 303–306. <https://doi.org/10.1002/anie.200704489>.
- (25) Dorcier, A.; Dyson, P. J.; McIndoe, J. S. Analysis of Coordination and Organometallic Compounds Using Photoionisation Mass Spectrometric Techniques. *Eur. J. Inorg. Chem.* **2003**, *2003* (24), 4294–4297. <https://doi.org/10.1002/ejic.200300459>.
- (26) Meurer, E. C.; Santos, L. S.; Pilli, R. A.; Eberlin, M. N. Probing the Mechanism of the Petasis Olefination Reaction by Atmospheric Pressure Chemical Ionization Mass and Tandem Mass Spectrometry. *Org. Lett.* **2003**, *5* (9), 1391–1394. <https://doi.org/10.1021/ol027439b>.
- (27) Santos, L. S.; Knaack, L.; Metzger, J. O. Investigation of Chemical Reactions in Solution Using API-MS. *Int. J. Mass Spectrom.* **2005**, *246* (1), 84–104. <https://doi.org/10.1016/j.ijms.2005.08.016>.
- (28) Jirásko, R.; Holčapek, M. Structural Analysis of Organometallic Compounds with Soft Ionization Mass Spectrometry. *Mass Spectrom. Rev.* **2011**, *30* (6), 1013–1036. <https://doi.org/10.1002/mas.20309>.
- (29) Collins, S.; Linnolahti, M.; Zamora, M. G.; Zijlstra, H. S.; Rodríguez Hernández, M. T.; Perez-Camacho, O. Activation of Cp₂ZrX₂ (X = Me, Cl) by Methylaluminumoxane As Studied by Electrospray Ionization Mass Spectrometry: Relationship to Polymerization Catalysis. *Macromolecules* **2017**, *50* (22), 8871–8884. <https://doi.org/10.1021/acs.macromol.7b00933>.
- (30) Henderson, M. A.; Trefz, T. K.; Collins, S.; Wang, M. Y.; McIndoe, J. S. Characterization of Isobutylaluminumoxanes by Electrospray Ionization Mass Spectrometry. *Organometallics* **2013**, *32* (7), 2079–2083. <https://doi.org/10.1021/om3010977>.
- (31) Rouholahnejad, F.; Mathis, D.; Chen, P. Narrowly Distributed Polyethylene via Reversible Chain Transfer to Aluminum by a Sterically Hindered Zirconocene/MAO. *Organometallics* **2010**, *29* (2), 294–302. <https://doi.org/10.1021/om900238k>.
- (32) Moscato, B.; Landis, C. Are Carbodiimide-Quenched Polyethylene Distributions Representative of Bulk Polymer Samples? Analysis of Metallocene-Catalyzed Ethylenepolymerization by ESI-MS, MALDI, GPC and NMR. *Chem. Commun.* **2008**, No. 44, 5785–5787. <https://doi.org/10.1039/B812133H>.
- (33) Penafiel, J.; Hesketh, A. V.; Granot, O.; Scott McIndoe, J. Electron Ionization Mass Spectrometric Analysis of Air- and Moisture-Sensitive Organometallic Compounds. *Dalton Trans.* **2016**, *45* (39), 15552–15556. <https://doi.org/10.1039/C6DT03020C>.
- (34) T. Chavagnan, A. Naim, M. Hubert, V. Cirriez, A. Welle, C. Afonso, P. Giusti, A. Vantomme, E. Kirillov, J.-F. Carpentier, n.d., (To Be Submitted Very Soon).
- (35) Giles, K.; Williams, J. P.; Campuzano, I. Enhancements in Travelling Wave Ion Mobility Resolution. *Rapid Communications in Mass Spectrometry* **2011**, *25* (11), 1559–1566. <https://doi.org/10.1002/rem.5013>.
- (36) McEwen, C. N.; McKay, R. G.; Larsen, B. S. Analysis of Solids, Liquids, and Biological Tissues Using Solids Probe Introduction at Atmospheric Pressure on Commercial LC/MS Instruments. *Anal. Chem.* **2005**, *77* (23), 7826–7831. <https://doi.org/10.1021/ac051470k>.
- (37) McEwen, C.; Gutteridge, S. Analysis of the Inhibition of the Ergosterol Pathway in Fungi Using the Atmospheric Solids Analysis Probe (ASAP) Method. *J. Am. Soc. Mass Spectrom.* **2007**, *18* (7), 1274–1278. <https://doi.org/10.1016/j.jasms.2007.03.032>.
- (38) Petucci, C.; Diffendal, J. Atmospheric Solids Analysis Probe: A Rapid Ionization Technique for Small Molecule Drugs. *J. Mass Spectrom.* **2008**, *43* (11), 1565–1568. <https://doi.org/10.1002/jms.1424>.
- (39) Jagerdeo, E.; Clark, J. A.; Leibowitz, J. N.; Reda, L. J. Rapid Analysis of Forensic Samples Using an Atmospheric Solid Analysis Probe Interfaced to a Linear Ion Trap Mass Spectrometer. *Rapid Commun. Mass Spectrom.* **2015**, *29* (2), 205–212. <https://doi.org/10.1002/rem.7097>.
- (40) Smith, M. J. P.; Cameron, N. R.; Mosely, J. A. Evaluating Atmospheric Pressure Solids Analysis Probe (ASAP) Mass Spectrometry for the Analysis of Low Molecular Weight Synthetic Polymers. *Analyst* **2012**, *137* (19), 4524–4530. <https://doi.org/10.1039/C2AN35556F>.
- (41) Barrère, C.; Maire, F.; Afonso, C.; Giusti, P. Atmospheric Solid Analysis Probe–Ion Mobility Mass Spectrometry of Polypropylene. *Anal. Chem.* **2012**, *84* (21), 9349–9354. <https://doi.org/10.1021/ac302109q>.
- (42) Farenc, M.; Corilo, Y. E.; Lalli, P. M.; Riches, E.; Rodgers, R. P.; Afonso, C.; Giusti, P. Comparison of Atmospheric Pressure Ionization for the Analysis of Heavy Petroleum Fractions with Ion Mobility–Mass Spectrometry. *Energy Fuels* **2016**, *30* (11), 8896–8903. <https://doi.org/10.1021/acs.energyfuels.6b01191>.
- (43) Mosely, J. A.; Stokes, P.; Parker, D.; Dyer, P. W.; Messinis, A. M. Analysis of Air-, Moisture- and Solvent-Sensitive Chemical Compounds by Mass Spectrometry Using an Inert Atmospheric Pressure Solids Analysis Probe. *Eur J Mass Spectrom (Chichester)* **2017**, *24* (1), 74–80. <https://doi.org/10.1177/1469066717732286>.
- (44) Constantine, S. P.; Cardin, D. J.; Bollen, B. G. Electron Impact and Chemical Ionisation Mass Spectrometry in the Characterisation of Organometallic Silicon, Germanium, Tin and Lead Compounds. *Rapid Communications in Mass Spectrometry* **2000**, *14* (5), 329–334. [https://doi.org/10.1002/\(SICI\)1097-0231\(20000315\)14:5<329::AID-RCM853>3.0.CO;2-T](https://doi.org/10.1002/(SICI)1097-0231(20000315)14:5<329::AID-RCM853>3.0.CO;2-T).
- (45) Farenc, M.; Assumani, B.; Afonso, C.; Giusti, P. Analysis of Metallocene Using Atmospheric Pressure Chemical Ionization and Atmospheric Pressure Photo Ionization Coupled with Ion Mobility Mass Spectrometry. *In Proceedings of the 64th ASMS conference on Mass Spectrometry and Allied Topics, San Antonio, TX, USA, 5–9 June 2016.* **2016**.
- (46) Dossmann (Soldi-Lose), H.; Afonso, C.; Lesage, D.; Tabet, J.-C.; Uggerud, E. Formation and Characterization of Gaseous Adducts of Carbon Dioxide to Magnesium, (CO₂)MgX– (X=OH, Cl, Br). *Angew. Chem. Int. Ed.* **2012**, *51* (28), 6938–6941. <https://doi.org/10.1002/anie.201108477>.
- (47) Dossmann, H.; Afonso, C.; Tabet, J.-C.; Uggerud, E. Unimolecular Dissociation Characteristics of Cationic Complexes between Nicotinic Acid and Cu(II) and Ni(II). *Int. J. Mass Spectrom.* **2013**, *354–355*, 165–174. <https://doi.org/10.1016/j.ijms.2013.05.030>.